



Clinical and surgical outcomes of robot-assisted versus conventional total hip arthroplasty: a systematic overview of meta-analyses

Nanne Kort¹

Patrick Stirling²

Peter Pilot³

Jacobus Hendrik Müller²

- Robot-assisted total hip arthroplasty (THA), in comparison to conventional THA, improves radiographic outcomes, but it remains unclear whether it alters complication rates, clinical and functional outcomes, and implant survival.
- The purpose of this systematic overview was to summarize the findings of the most recent meta-analyses that compare clinical and surgical outcomes of robot-assisted versus conventional THA.
- Two readers independently conducted an electronic literature search, screening and data extraction from five electronic databases. Inclusion criteria were: meta-analyses evaluating robot-assisted versus conventional THA in terms of radiographic outcomes, clinical and functional scores, and complications and revision rates. The literature search returned 67 records, of which 14 were duplicates and 49 were excluded, leaving three meta-analyses published within the past two years for data extraction and analysis.
- The present overview of meta-analyses suggests that, compared to conventional THA ($n = 3011$), robot-assisted THA ($n = 1813$) improves component placement and reduces intraoperative complications. The overview also affirms that robot-assisted THA could extend surgery by 20 minutes, and increases risks of postoperative heterotopic ossification, dislocation, and revision. None of the meta-analyses found significant differences in clinical or functional scores between robot-assisted and conventional THA.
- Future studies and reviews should make a clear distinction between active and semi-active robotic assistance, address technology maturity, and describe the experience of surgeons with robotic assistance.

Keywords: clinical and radiographic outcomes; robotic surgical procedures; total hip arthroplasty

Cite this article: *EFORT Open Rev* 2021;6:1157-1165.

DOI: 10.1302/2058-5241.6.200121

Introduction

In an attempt to improve accuracy and consistency of implant placement during total hip arthroplasty (THA), multiple navigation technologies have been introduced over the past three decades,¹ which can be broadly characterized as computer-assisted navigation systems, or robot-assisted systems.² Robotic systems, which are utilized across many surgical subspecialties,³ can be classified as either active systems, which work autonomously to perform the planned bone resections, or semi-active systems, which provide full control to the surgeon with live intraoperative feedback to limit deviation from the preoperative surgical plan.⁴ Although robotic assistance in THA improves precision and accuracy,⁵ it remains unclear whether it alters complication rates, clinical and functional outcomes, and implant survival.⁶

Over the last three years, numerous meta-analyses pooled data from published studies that compared outcomes of robot-assisted versus conventional THA. To the authors' knowledge, there is no published overview of these meta-analyses to summarize the latest evidence in terms of the effect of robot-assisted THA on rates of complications, clinical and functional outcomes, or implant survival. The purpose of this overview was therefore to summarize the findings of the most recent meta-analyses on the efficacy of robot-assisted versus conventional THA, and highlight any differences in surgical and clinical outcomes. This overview is expected to highlight gaps in the literature and help decision-makers justify clinical and economic benefits of robotic assistance.⁷

Material and methods

The protocol for this overview of systematic reviews and meta-analyses, including the search strategy and

proposed methodology, was registered with PROSPERO (CRD42020181669).

Search strategy

The authors conducted an electronic literature search using Allied & Complementary Medicine™, Embase®, MEDLINE®, Web of Science, Cochrane Database of Systematic Reviews on 11 February 2020. Key words used to develop search strategy were (“hip” OR “knee”) AND (“arthroplasty” OR “replacement”) AND (“robot” OR “robot*”) AND “meta-analysis” (see full search strategy in PROSPERO registration). While the original search strategy included both hip and knee arthroplasty, it was subsequently resolved that only results regarding primary THA would be included.

Two reviewers (JHM and KJC) independently performed the literature search described. Grey literature regarding robot-assisted THA was searched and an expert in the field (NK) consulted for other relevant publications not identified in the electronic search. Review registries were checked for ongoing reviews on the subject. Disagreements between reviewers were discussed and resolved by consensus.

Study selection and data extraction

Titles and abstracts of the studies were screened independently by two reviewers (JHM and KJC) to determine relevance according to the inclusion and exclusion criteria presented below.

Inclusion criteria

Original meta-analyses that:

- reported on studies evaluating robot-assisted, both active or semi-active, compared to conventional primary THA for any indication.
- presented results in terms of:
 - radiographic outcomes (such as limb and joint alignment, component placement).
 - clinical scores (such as Harris Hip Score (HHS), Western Ontario and McMaster Universities Osteoarthritis Index (WOMAC)).
 - complication rates (intra- and postoperative complications).
 - implant survival or revision rates (such as Kaplan-Meier, Cumulative Incidence Function).

Exclusion criteria

Meta-analyses that:

- reported outcomes for robot-assisted surgery for other joints without separating data regarding THA.
- were written in languages other than English, to avoid translation errors.

Full-text articles were retrieved if the article passed the first eligibility screening or if the title or abstract provided insufficient information to establish eligibility. Disagreements in screening decisions between the reviewers were discussed and resolved by review and consensus. The reference lists of all selected publications were checked for relevant studies that may have been missed in the electronic search.

Data extraction and quality assessment

Two reviewers (JHM and KJC) extracted characteristics of meta-analyses independently including: year of publication, journal, number and type of studies included, countries in which included studies were performed, intervention and comparator details, number of patients included per intervention and comparator, follow-up period, type of robot used, pooled outcomes recorded by at least three studies. Pooled outcome data reported by the meta-analyses included reported effect size and statistical significance. Results of data extraction were compared and where discrepancies were found, consensus was reached through review and discussion between the reviewers.

The same two reviewers (JHM and KJC) assessed the methodological quality of eligible studies according to the 16 domains outlined by A Measurement Tool to Assess systematic Reviews (AMSTAR-2).⁸ Where there was disagreement between reviewers in their appraisal of study quality, consensus was achieved through review and discussion.

Interpretation of results

Methodological differences across meta-analyses made pooling or direct statistical comparison of results impossible. As a result, findings extracted from each meta-analysis were presented as reported and synthesized narratively, rather than normalized to a single comparable metric. Differences in outcomes were reported as weighted mean difference (WMD) or weighted odds ratio (WOR) and considered statistically significant if $p < 0.05$.

Results

Literature search

The electronic literature search returned 67 records, of which 14 were duplicates. A further 49 articles were excluded after reading their titles or abstracts (46 did not include THA; two did not include robotic assistance, and one was not written in English), and an additional article⁹ was excluded after reading its full text, as it included < 3 studies per outcome of interest, leaving a total of three meta-analyses eligible for quality assessment and data extraction (Fig. 1).^{2,10,11}




Fig. 1 Flowchart of the study selection procedure.

Note. THA, total hip arthroplasty; RA, robotic-assisted.

Characteristics of included studies

The three meta-analyses, all published within the past two years, assessed a total of 15 comparative studies reporting outcomes of 1813 hips that received robot-assisted THA and 3011 hips that received conventional THA. The majority of studies originated from the USA ($n = 7$),^{5,12–17} and the most frequently used system was the ROBODOC which provides active assistance (THINK Surgical, Inc., Fremont, CA, $n = 8$ studies).^{12,18–24} Of the 15 studies, five,^{19,21,23–25} were included by all three meta-analyses, two^{12,20} were included by both Han et al² and Karunaratne et al,¹⁰ two^{16,17} were included by both Han et al² and Chen et al,¹¹ five^{5,14,15,18,22} were included only by Han et al,² and one¹³ was included only by Karunaratne et al¹⁰ (Table 1). It is worth noting that all three meta-analyses pooled results of older and possibly obsolete robotic systems with results of

newer generations and enhanced robotic systems, which may be a methodological flaw. While the meta-analyses did not distinguish between outcomes of old (such as ROBODOC (ORTHODOC) and Caspar) and new (such as Mako) systems, which makes it impossible to present their results separately, inspection of forest plots revealed no consistent differences in outcomes of old versus new systems.

Quality assessment of included meta-analyses

According to AMSTAR-2, methodological quality was ‘low’ for Karunaratne et al¹⁰ due to weakness in a critical domain, and ‘critically low’ for the remaining two studies^{2,11} due to weaknesses in two or more critical domains (Table 2, Fig. 2). All three meta-analyses failed to apply appropriate methods for data synthesis; neither Han et al² nor Chen et al¹¹ prospectively published their review protocols in

Table 1. Characteristics of included studies

	Han et al, 2019 ²	Karunaratne et al, 2019 ¹⁰	Chen et al, 2018 ¹¹
Journal	<i>Int J Med Robot Comp</i>	<i>Int Orthop</i>	<i>Postgrad Med J</i>
Population	THA	THA	THA
Intervention (robotic-assisted, hips)	817	474	522
Comparator (conventional, hips)	1536	481	994
Outcomes reported by ≥ 3 studies			
Radiographic outcomes	yes		yes
Clinical scores	yes	yes	yes
Complication and revision rates	yes		yes
Operation time	yes		yes
Follow-up (months, range)	(0–168)	(18–60)	(0–60)
Studies assessed			
Total (unique inclusions)	14 (5)	8 (1)	7 (0)
RCT	5	4	2
Cohort	7	1	1
Case-control	8	3	4
Robots			
ROBODOC	8	6	4
CASPAR	1	1	1
MAKO	5	1	2
Countries			
USA	6	2	2
Japan	4	3	3
Germany	3	2	1
Korea	1	1	1

Note. THA, total hip arthroplasty; RCT, randomized controlled trial.

Table 2. Evaluation of the quality of meta-analyses on RA THA using AMSTAR-2

First author, year	Han et al, 2019 ²	Karunaratne et al, 2019 ¹⁰	Chen et al, 2018 ¹¹
Intervention	THA	THA & TKA	THA
1. Research questions and criteria included PICO	N	Y	Y
2. Published review protocol prior (c)	N	Y	N
3. Explained study design inclusion criteria	N	N	Y
4. Comprehensive literature search strategy (c)	P	Y	P
5. Performed study selection in duplicate	Y	Y	Y
6. Performed data extraction in duplicate	Y	Y	Y
7. Excluded studies listed and justified (c)	N	Y	N
8. Included studies described in adequate detail	P	P	P
9. Included studies assessed for RoB (c)	Y	Y	P
10. Reported sources of funding for studies	N	N	N
11. Appropriate methods for data synthesis (c)	N	N	N
12. Assessed impact of RoB in each study	Y	Y	N
13. Considered RoB when interpreting results (c)	N	Y	N
14. Observed heterogeneity & impact explained	N	Y	N
15. Investigated publication bias (c)	Y	Y	N
16. Reported own conflict of interests & funding	Y	Y	Y
Number of critical weaknesses	4	1	5
Result (AMSTAR-2)	Critically low	Low	Critically low

Note. THA, total hip arthroplasty; TKA, total knee arthroplasty; RA, robotic-assisted; AMSTAR-2, A Measurement Tool to Assess systematic Reviews; PICO, Population Intervention Comparator Outcome; RoB, risk of bias; c, critical.

advance, listed or justified excluded studies, or considered risk of bias when interpreting their results.

Radiographic outcomes

Chen et al¹¹ reported on radiographic outcomes, which could not be considered because they had fewer than three

clinical studies on each outcome. Han et al² reported on radiographic outcomes, including acetabular cup inclination, cup anteversion, stem alignment, cup safe zones (Lewinnek and Callanan) and leg length discrepancy (> 3 or > 10 mm). They found that robot-assisted THA improved both cup inclination (WMD, 2.47°; p = 0.03) and stem



Fig. 2 Critical AMSTAR-2 domains for the assessment of the meta-analyses.

Note. AMSTAR-2, A MeaSurement Tool to Assess systematic Reviews; PICO, Population Intervention Comparator Outcome; RoB, risk of bias; c, critical.

alignment (WMD, 0.4°; p = 0.02), as well as positioning within the Lewinnek safe zone (WOR, 11.05; p < 0.001) and the Callanan safe zone (WOR, 7.63; p < 0.001) (Table 3).

Functional outcomes

Han et al² reported weighted HSS, Postel-Merle d’Aubigné (PMA) and pooled different scores (HSS, PMA and Japanese Orthopedic Association (JOA)). Karunaratne et al¹⁰ reported weighted PMA and pooled HHS and modified HHS together. Chen et al¹¹ did not report any weighted scores, but pooled different scores (HSS, PMA and JOA). None of the meta-analyses found statistically significant differences in clinical scores between robot-assisted and conventional THA (Table 3).

Complications and survival

Both Han et al² and Chen et al¹¹ found that robot-assisted THA decreased intraoperative complications (respectively: WOR, 0.32; p = 0.006 and WOR, 0.12; p < 0.001). Chen et al¹¹ found that robot-assisted THA decreased overall complications (WOR, 0.42; p = 0.03), whereas Han et al² found no significant difference. Han et al² reported that

robot-assisted THA increased dislocation (WOR, 2.28; p = 0.02) and revisions (WOR, 2.88; p = 0.03), and Chen et al¹¹ likewise reported that robot-assisted THA increased heterotopic ossification (WOR, 1.94; p = 0.04) (Table 3).

Operation time

Both Han et al² and Chen et al¹¹ found that robot-assisted THA extends operation time by about 20 minutes. Han et al² found a statistically significant difference (WMD, 20.72 minutes; p = 0.002), while Chen et al¹¹ did not (WMD, 23.21 minutes) (Table 3).

Conclusions of meta-analyses

All three meta-analyses concluded that postoperative clinical results were equivalent, with both Chen et al¹¹ and Karunaratne et al¹⁰ calling for further studies to ascertain long-term outcomes. Both Chen et al¹¹ and Han et al² further concluded that while robot-assisted THA requires longer operation times, it incurs fewer intraoperative complications and better radiographic outcomes. Chen et al¹¹ also concluded that robot-assisted THA increases likelihood of heterotopic ossification, while Han et al²

Table 3. All reported outcomes of THA using robotic assistance and conventional instrumentation

	Unit	Han et al, 2019 ²				Karunaratne et al, 2019 ¹⁰				Chen et al, 2018 ¹¹			
		n*	Effect size	p-value	Favours	n*	Effect size	p-value	Favours	n*	Effect size	p-value	Favours
Radiographic outcomes													
Cup inclination (degrees)	WMD	4	-2.47	0.003	RA								
Cup anteversion (degrees)	WMD	4	-1.63	0.600									
Stem alignment (degrees)	WMD	6	-0.40	0.020	RA								
Cup safe zone Lewinnek	WOR	4	11.05	< 0.001	CI								
Cup safe zone Callanan	WOR	4	7.63	< 0.001	CI								
LLD (> 3 or > 10 mm)	WOR	4	0.74	0.280									
Clinical scores													
Pooled HHS, PMA & JOA score	WMD	10	0.01	0.970					3	0.09	0.380		
Pooled mHHS and HHS	WMD					4	-2.90	n.r.					
PMA score	WMD	4	0.06	0.860		4	-1.25	n.r.					
HHS	WMD	4	0.04	0.980									
Complications and revision													
Intraoperative complication	WOR	9	0.32	0.006	RA				5	0.12	< 0.0001	RA	
Nerve palsy	WOR	3	4.47	0.110									
Thigh pain	WOR	3	0.32	0.030	RA								
Heterotopic ossification	WOR	4	1.44	0.290					3	1.94	0.040	CI	
Dislocation	WOR	6	2.28	0.020	CI								
Total complications	WOR	7	0.83	0.480					5	0.43	0.030	RA	
Revision rate	WOR	3	2.88	0.030	CI								
Operation time (minutes)	WMD	8	20.72	0.002	CI				3	23.21	0.090		

*Number of studies assessing an outcome.

Note. THA, total hip arthroplasty; RA, robotic-assisted; CI, conventional instrumentation; WMD, weighted mean difference; WOR, weighted odds ratio; HHS, Harris Hip Score; PMA, Postel-Merle d'Aubigné; JOA, Japanese Orthopaedic Association; mHHA, modified Harris Hip Score; LLD, leg length discrepancy; n.r., not reported.

concluded that it is associated with a higher incidence of dislocations and revisions.

Discussion

The present overview of meta-analyses suggests that, compared to conventional THA, robot-assisted THA grants more accurate cup inclination and stem alignment, higher likelihood of component placement within safe zones, and fewer intraoperative complications. The overview also affirms that robot-assisted THA extends operation times by about 20 minutes, and increases risks of postoperative heterotopic ossification, dislocation, and revision. None of the meta-analyses found significant differences in clinical or functional scores between robot-assisted and conventional THA.

The two meta-analyses^{2,11} that reported on radiographic outcomes found that, compared to conventional THA, robot-assisted THA enabled more accurate and reproducible acetabular cup placement within the Lewinnek safe zone²⁶ and the Callanan safe zone.²⁷ The validity of both safe zones has been challenged,^{28–30} because subluxations and dislocations have also been observed for cups that were placed within the safe zones. As a result, several additional safe zones have been proposed that show an improved accuracy of component positioning,^{29–31} but these were not used in the studies assessed by the meta-analyses. Moreover, a recent systematic review on acetabular cup positioning and risk of dislocation suggested that

it is difficult to draw any conclusions regarding definitive target zones for cup positioning due to high heterogeneity among studies with inconsistent measurement techniques and different surgical approaches.³² The authors therefore believe that ideal cup placement should be determined considering spino-pelvic parameters, such as pelvic tilt and functional anteversion, which could be facilitated by a robotic system.^{33–35}

The meta-analyses revealed more accurate stem placement with robot-assisted THA, but there remains inconsistency in standards for classification of stem alignment.¹¹ Leg length discrepancy (LLD) remains one of the most common causes of patient dissatisfaction after THA, though there is no consensus as to whether the cut-off should be 3 mm, 5 mm or 10 mm.^{2,11} The three meta-analyses found no statistically significant differences in LLD between robot-assisted and conventional THA, either in terms of absolute difference or proportion of outliers.^{2,11} Robotic systems provide an accurate way to assess LLD that may help surgeons make intraoperative adjustments and/or improve their preoperative planning or component positioning.

Based on the findings of the current overview, robotic assistance has no added benefit in terms of clinical and functional scores at 5 to 14 years.^{2,10,11} It should be noted, however, that the use of different scoring systems across studies complicates evaluation of any pooled results. Han et al² noted that in a study by Bargar et al,¹⁸ robotic assistance yielded significantly better pain scores (Health

Status Questionnaire and Harris Pain Scores) as well as WOMAC scores at a mean follow-up of 14 years (robot-assisted THA, 13.9±2.7 years; conventional THA, 14.2±4.7 years).

In the meta-analysis by Chen et al¹¹ the rates of infection, nerve palsy and deep vein thrombosis were comparable between robot-assisted and conventional THA. Han et al² revealed significantly higher dislocation and revision rates with robotic assistance. It is worth noting, however, that studies published after 2003 observed lower dislocation rates following robot-assisted THA.²⁰ This decrease might be attributable to the inclusion of five studies^{19,21,23–25} that followed a posterolateral approach, which provides better retraction of the gluteus medius and minimus muscles, thereby granting improved access for robotic milling and avoiding injury to the abductor tendon and greater trochanter.²³ It is noteworthy that studies evaluating active robot-assisted THA reported outcomes at 1.5 to 14 years,^{12,18–25} whereas studies evaluating semi-active robot-assisted THA reported outcomes at only 0 to 2 years.^{5,13–17} The long-term outcomes of semi-active robot-assisted THA are therefore yet to be determined.

Both of the meta-analyses that assessed operation time indicated that robot-assisted THA took longer than conventional THA, possibly because robotic systems require registration or placement of positioning pins, as well as the learning curve for new users. The latter has not been addressed in the meta-analyses which did not consider the level of experience of the surgeons. There are few reports on the learning curve of robot-assisted THA.² One study observed a significant learning curve, with operation time decreasing from 79.8 minutes (1st to 35th case) to 69.4 minutes (71st to 105th cases),³⁶ whereas another study found surgeons were able to grasp the technology after only 10 procedures.⁵ A third study compared one surgeon's experience switching from conventional to robot-assisted THA, and found that over the course of 100 surgeries, it took 14 surgeries to become 'proficient', beyond which there were no significant differences in operation time or HHS.³⁷

The findings of this overview of meta-analyses should be interpreted with the following considerations and limitations in mind. First, only three meta-analyses fulfilled the inclusion criteria, and their quality was either 'low' or 'critically low'. Moreover, all three meta-analyses included nine case-control studies and one cohort study in addition to five randomized controlled trials. Second, only one meta-analysis¹⁰ differentiated between active and semi-active assistance, whereas results from both systems were pooled in the other two meta-analyses.^{2,11} Moreover, all three meta-analyses pooled results of older and possibly obsolete robotic systems with results of newer generations and enhanced robotic systems. This may be problematic as blending results across different robotic

assistance techniques and generations may invalidate the data syntheses performed. Third, there was heterogeneity in terms of surgical approaches, and it is impossible to differentiate the effect of surgical approach from that of robotic assistance. Fourth, it is impossible to account for the effect of learning curves and experience in the included meta-analyses. Fifth, only the Lewinnek et al²⁶ and Callanan et al²⁷ safe zones for acetabular component positioning were assessed, while newer safe zones were not accounted for. Safe zones enable quantitative assessments of how well surgeons followed their preoperative plans, and hence how to improve their techniques and targets for future operations. Sixth, 'human error' remains a major weakness in THA,⁶ since it is impossible to implant perfectly positioned components in every patient with their varying biological environments, diverse anatomy, and pathology. It is unknown whether and in how many cases surgeons might have diverted from the preoperative plan, and how this affected the reported outcomes. Seventh, technology has evolved greatly over the last two decades and is still evolving very fast.⁷ Therefore the question arises whether data can be pooled for technologies of different generations and working methods.

Conclusion

The present overview of meta-analyses suggests that robot-assisted THA could improve the accuracy of component positioning and reduce intraoperative complications. The overview also affirms that robot-assisted THA extends surgery by 20 minutes, and increases risks of postoperative heterotopic ossification, dislocation, and revision. None of the meta-analyses found significant differences in clinical or functional scores between robot-assisted and conventional THA. Future studies and reviews should make a clear distinction between active and semi-active robotic assistance, address technology maturity, and consider surgeon experience.

AUTHOR INFORMATION

¹CortoClinics, Nederweert, Netherlands.

²ReSurg SA, Nyon, Switzerland.

³IMUKA, Roosteren, The Netherlands.

Correspondence should be sent to: Jacobus Hendrik Müller, ReSurg SA, Rue Saint-Jean 22, 1260 Nyon Switzerland.
Email: team@resurg.com

ACKNOWLEDGEMENTS

The authors are grateful for the contributions of Kyung-Jin Cho who assisted with the electronic literature collection and screening process. The authors are grateful to Mo Saffarini for his assistance with manuscript preparation.

ICMJE CONFLICT OF INTEREST STATEMENT

NK reports being a consultant for BodyCad, Stryker, Zimmer-Biomet, Bioventus, outside the submitted work. PS reports no conflicts of interest. PP reports former employment at Zimmer Biomet. JHM reports no conflicts of interest. The review and manuscript preparation were funded by Stichting IMA.

FUNDING STATEMENT

The author or one or more of the authors have received or will receive benefits for personal or professional use from a commercial party related directly or indirectly to the subject of this article.

OPEN ACCESS

© 2021 The author(s)

This article is distributed under the terms of the Creative Commons Attribution-Non Commercial 4.0 International (CC BY-NC 4.0) licence (<https://creativecommons.org/licenses/by-nc/4.0/>) which permits non-commercial use, reproduction and distribution of the work without further permission provided the original work is attributed.

REFERENCES

- Renner L, Janz V, Perka C, Wassilew GI. What do we get from navigation in primary THA? *EFORT Open Rev* 2017;1:205–210.
- Han PF, Chen CL, Zhang ZL, et al. Robotics-assisted versus conventional manual approaches for total hip arthroplasty: a systematic review and meta-analysis of comparative studies. *Int J Med Robot* 2019;15:e1990.
- D'Souza M, Gendreau J, Feng A, Kim LH, Ho AL, Veeravagu A. Robotic-assisted spine surgery: history, efficacy, cost, and future trends. *Robot Surg* 2019;6:9–23.
- Kayani B, Konan S, Ayuob A, Onochie E, Al-Jabri T, Haddad FS. Robotic technology in total knee arthroplasty: a systematic review. *EFORT Open Rev* 2019;4:611–617.
- Kamara E, Robinson J, Bas MA, Rodriguez JA, Hepinstall MS. Adoption of robotic vs fluoroscopic guidance in total hip arthroplasty: is acetabular positioning improved in the learning curve? *J Arthroplasty* 2017;32:125–130.
- Subramanian P, Wainwright TW, Bahadori S, Middleton RG. A review of the evolution of robotic-assisted total hip arthroplasty. *Hip Int* 2019;29:232–238.
- Zagra L. Advances in hip arthroplasty surgery: what is justified? *EFORT Open Rev* 2017;2:171–178.
- Shea BJ, Reeves BC, Wells G, et al. AMSTAR 2: a critical appraisal tool for systematic reviews that include randomised or non-randomised studies of healthcare interventions, or both. *BMJ* 2017;358:j4008.
- Tan A, Ashrafian H, Scott AJ, et al. Robotic surgery: disruptive innovation or unfulfilled promise? A systematic review and meta-analysis of the first 30 years. *Surg Endosc* 2016;30:4330–4352.
- Karunaratne S, Duan M, Pappas E, et al. The effectiveness of robotic hip and knee arthroplasty on patient-reported outcomes: a systematic review and meta-analysis. *Int Orthop* 2019;43:1283–1295.
- Chen X, Xiong J, Wang P, et al. Robotic-assisted compared with conventional total hip arthroplasty: systematic review and meta-analysis. *Postgrad Med J* 2018;94:335–341.
- Bargar WL, Bauer A, Börner M. Primary and revision total hip replacement using the Robodoc system. *Clin Orthop Relat Res* 1998;354:82–91.
- Bukowski BR, Anderson P, Khlopas A, Chughtai M, Mont MA, Ilgen RL II. Improved functional outcomes with robotic compared with manual total hip arthroplasty. *Surg Technol Int* 2016;29:303–308.
- Tsai TY, Dimitriou D, Li JS, Kwon YM. Does haptic robot-assisted total hip arthroplasty better restore native acetabular and femoral anatomy? *Int J Med Robot* 2016;12:288–295.
- El Bitar YF, Stone JC, Jackson TJ, Lindner D, Stake CE, Domb BG. Leg-length discrepancy after total hip arthroplasty: comparison of robot-assisted posterior, fluoroscopy-guided anterior, and conventional posterior approaches. *Am J Orthop* 2015;44:265–269.
- Domb BG, El Bitar YF, Sadik AY, Stake CE, Botser IB. Comparison of robotic-assisted and conventional acetabular cup placement in THA: a matched-pair controlled study. *Clin Orthop Relat Res* 2014;472:329–336.
- Domb BG, Redmond JM, Louis SS, et al. Accuracy of component positioning in 1980 total hip arthroplasties: a comparative analysis by surgical technique and mode of guidance. *J Arthroplasty* 2015;30:2208–2218.
- Bargar WL, Parise CA, Hankins A, Marlen NA, Campanelli V, Netravali NA. Fourteen year follow-up of randomized clinical trials of active robotic-assisted total hip arthroplasty. *J Arthroplasty* 2018;33:810–814.
- Hananouchi T, Sugano N, Nishii T, et al. Effect of robotic milling on periprosthetic bone remodeling. *J Orthop Res* 2007;25:1062–1069.
- Honl M, Dierk O, Gauck C, et al. Comparison of robotic-assisted and manual implantation of a primary total hip replacement: a prospective study. *J Bone Joint Surg [Am]* 2003;85-A:1470–1478.
- Lim SJ, Ko KR, Park CW, Moon YW, Park YS. Robot-assisted primary cementless total hip arthroplasty with a short femoral stem: a prospective randomized short-term outcome study. *Comput Aided Surg* 2015;20:41–46.
- Nakamura N, Sugano N, Nishii T, Miki H, Kakimoto A, Yamamura M. Robot-assisted primary cementless total hip arthroplasty using surface registration techniques: a short-term clinical report. *Int J CARS* 2009;4:157–162.
- Nakamura N, Sugano N, Nishii T, Kakimoto A, Miki H. A comparison between robotic-assisted and manual implantation of cementless total hip arthroplasty. *Clin Orthop Relat Res* 2010;468:1072–1081.
- Nishihara S, Sugano N, Nishii T, Miki H, Nakamura N, Yoshikawa H. Comparison between hand rasping and robotic milling for stem implantation in cementless total hip arthroplasty. *J Arthroplasty* 2006;21:957–966.
- Siebel T, Käfer W. [Clinical outcome following robotic assisted versus conventional total hip arthroplasty: a controlled and prospective study of seventy-one patients]. *Z Orthop Ihre Grenzgeb* 2005;143:391–398.
- Lewinnek GE, Lewis JL, Tarr R, Compere CL, Zimmerman JR. Dislocations after total hip-replacement arthroplasties. *J Bone Joint Surg [Am]* 1978;60-A:217–220.
- Callanan MC, Jarrett B, Bragdon CR, et al. The John Charnley Award: risk factors for cup malpositioning: quality improvement through a joint registry at a tertiary hospital. *Clin Orthop Relat Res* 2011;469:319–329.
- Abdel MP, von Roth P, Jennings MT, Hanssen AD, Pagnano MW. What safe zone? The vast majority of dislocated THAs are within the Lewinnek safe zone for acetabular component position. *Clin Orthop Relat Res* 2016;474:386–391.
- Reina N, Putman S, Desmarchelier R, et al; SFHG. Can a target zone safer than Lewinnek's safe zone be defined to prevent instability of total hip arthroplasties? Case-control study of 56 dislocated THA and 93 matched controls. *Orthop Traumatol Surg Res* 2017;103:657–661.
- Tezuka T, Heckmann ND, Bodner RJ, Dorr LD. Functional safe zone is superior to the Lewinnek safe zone for total hip arthroplasty: why the Lewinnek safe zone is not always predictive of stability. *J Arthroplasty* 2019;34:448–458.

31. **Barrack RL, Krempec JA, Clohisy JC, et al.** Accuracy of acetabular component position in hip arthroplasty. *J Bone Joint Surg Am* 2013;95:1760–1768.
32. **Seagrave KG, Troelsen A, Malchau H, Husted H, Gromov K.** Acetabular cup position and risk of dislocation in primary total hip arthroplasty. *Acta Orthop* 2017;88:10–17.
33. **Behery OA, Poultsides L, Vigdorich JM.** Modern imaging in planning a personalized hip replacement and evaluating the spino-pelvic relationship in prosthetic instability. In: Rivière C, Vendittoli PA, eds. *Personalized hip and knee joint replacement*. Cham: Springer, 2020:143–156.
34. **Perticarini L, Rossi SMP, Benazzo F.** Unstable total hip replacement: why? Clinical and radiological aspects. *Hip Int* 2020;30:37–41.
35. **Sultan AA, Khlopas A, Piuze NS, Chughtai M, Sodhi N, Mont MA.** The impact of spino-pelvic alignment on total hip arthroplasty outcomes: a critical analysis of current evidence. *J Arthroplasty* 2018;33:1606–1616.
36. **Redmond JM, Gupta A, Hammarstedt JE, Petrakos AE, Finch NA, Domb BG.** The learning curve associated with robotic-assisted total hip arthroplasty. *J Arthroplasty* 2015;30:50–54.
37. **Kong X, Yang M, Jerabek S, Zhang G, Chen J, Chai W.** A retrospective study comparing a single surgeon's experience on manual versus robot-assisted total hip arthroplasty after the learning curve of the latter procedure: a cohort study. *Int J Surg* 2020;77:174–180.